

Table of assigned values Waste acc to landfill directive (eluat ions) - AB09

1<sup>st</sup> Edition, 23.10.2020

## Table of assigned values

Parameter	Sample	Unit	Assigned value ±	U (k=2)	Criterion	Criterion [%]
Evaporation residue	AB09	mg/l	325 ±	7.82	17.2	5.3
Chloride	AB09	mg/l	27.8 ±	0.343	1.39	5
El. conductivity (25°C)	AB09	mS/m	53.3 ±	0.324	0.8	1.5
Fluorid	AB09	mg/l	0.523 ±	0.0258	0.0627	12
NH4 (as N)	AB09	mg/l	1.26 ±	0.0378	0.091	7.2
NO2 (as N)	AB09	mg/l	0.202 ±	0.00318	0.0101	5
NO3 (as N)	AB09	mg/l	3.31 ±	0.0693	0.166	5
pH-value	AB09		7.93 ±	0.0361	0.159	2
PO4 (as P)	AB09	mg/l	0.111 ±	0.0197	0.0387	35
Sulfate (as SO4)	AB09	mg/l	60.2 ±	1.06	3.01	5
TOC (as C)	AB09 -	mg/l	7.38 ±	0.359	0.886	12

### Legend:

Assigned value	Target value for proficiency assessment of the participants (3 significant digits)
U (k=2)	Expanded uncertainty (k=2) of the assigned value (3 significant digits)
Criterion	Specified value for the determination of the z-score in the given unit (3 significant digits)
Criterion [%]	Specified value for the determination of the z-score in % of the assigned value (2 significant digits)