

EVALUATION OF THE INTERLABORATORY COMPARISON TEST

Chlorinated hydrocarbons (CHC) and BTEX & C5-C10 – CBL04

Sample dispatch on 23rd October 2018

Edition 18th January 2019

Address: Umweltbundesamt GmbH
Spittelauer Lände 5
1090 Vienna/Austria

Contact: Dr. Sandra Kulcsar

Telephone: +43 (0) 1 31304 4334

E-mail: ringversuche@umweltbundesamt.at

Website: http://www.umweltbundesamt.at/en/interlaboratory_comparison/

Management:
Dipl.-Ing. Monika Denner

Table of contents

1	Description of the Interlaboratory comparison test: CHC and BTEX & C5-C10 – CBL04.....	4
1.1	Participants and time schedule.....	4
1.2	Sampling, sample material and distribution	4
1.3	Control testing	5
2	Evaluation	5
3	Representation and interpretation of measurement results.....	6
4	Explanatory notes	6
5	Annotations on tables and charts.....	7
5.1	Information and abbreviations in tables.....	7
5.2	Graphical presentation of results	9
6	Summary of results, after removal of outliers.....	11
7	Parameter oriented report.....	12
8	Laboratory oriented report.....	85

1 Description of the Interlaboratory comparison test: CHC and BTEX & C5-C10 – CBL04

1.1 Participants and time schedule

- Number of registrations: 22
- Number of submitted data records: 22
- Dispatch of samples: 23rd October 2018
- Closing date for submission of data: 20th November 2018

For the interlaboratory comparison test CBL04 the participants could participate in CL05 (CHC) and/or BL06 (BTEX & C5-C10).

To anonymize results, each laboratory was given a laboratory code on a random basis.

1.2 Sampling, sample material and distribution

Activated Orbo 32S-charcoal tubes (Supelco) were loaded using a certified calibration gas (Air Liquide). One tube was loaded with cis-1,2-Dichlorethene, trans-1,2-Dichlorethene, Trichloromethane, 1,1,1-Trichloroethane, Trichloroethene, Tetrachloromethane and Tetrachloroethene (CL05) and another tube was loaded with benzene, ethylbenzene o-, m- and p- xylene, toluene, n-Pentane, n-Hexane, n-Heptane, n-Octane, n-Nonane and n-Decane (BL06). In addition to CL05 and BL06, respectively, an unloaded activated charcoal tube was made available to determine the blank value. The tubes were loaded using a t-piece under pressure-less condition. The samples were prepared in two series (CL05 and BL06) using a pump with a continuous and defined flow. The flow of the pump was controlled before as well as after the loading of the tubes. The charcoal tubes were loaded on 19th October and 22nd October 2018 and dispatched on 23rd October 2018.

Each participant received (according to the order):

1 loaded charcoal tube for the interlaboratory comparison test CL05 and/or

1 loaded charcoal tube for the interlaboratory comparison test BL06

1 unloaded charcoal tube (blank)

1.3 Control testing

During sampling, aliquots of each sample were collected randomly for control testing. The parameters were determined by an external laboratory (accredited according to EN ISO/IEC 17025 standard).

In the parameter-oriented evaluation, the results of the control testing are given in the form of arithmetic means of the detected concentrations as control test value $\pm U$.

2 Evaluation

The analytical results had to be made available to the organiser not later than 20th November 2018. Any values received at a later date were not considered. A statistical evaluation of interlaboratory comparison data was only carried out if at least 6 valid results per parameter were available.

To evaluate the data, outliers were detected first by using the outlier test method according to Hampel. Values identified as conspicuous by this test method are marked specifically in the parameter-oriented evaluation.

Further evaluation was performed in accordance with DIN ISO 5725-2. Results $< LOQ$ or $< LOD$ are not taken into account for calculation.

The adjusted average value (after removal of outliers) for all submitted results was used as a basis for the calculation of recovery rates and z-scores.

z-Score

z-Scores were calculated on the basis of the following formula:

$$z - score = \frac{x_i - \bar{X}}{Criteria}$$

In this context,

x_i	is the measurement value of the participating laboratory.
\bar{X}	is the target value, normally the average value of the participants' results after removal of outliers; if this approach is not applicable, the target value is assigned according to the procedure given in section 4;
Criteria	is normally the reproducibility standard deviation (sR) calculated from the participants' results (after removal of outliers) in the relevant test round; if this approach is not applicable, the criteria is derived according to the procedure given in section 4

Interpretation of z-Scores in the parameter-oriented evaluation:

- $|z| < 2$ result: good
- $2 < |z| < 3$ result: questionable
- $|z| > 3$ result: not satisfactory

3 Representation and interpretation of measurement results

The parameter oriented report shows the measurement values including uncertainty, recovery rate, calculated z-Score and the outliers in tabular form. The results listed in the table are also represented graphically.

The laboratory oriented report shows the results of the individual laboratories, including the recovery rates and z-Scores.

An annotation of the tables and graphics is given in section 5.

4 Explanatory notes

As explained in section 2, the z-Score is normally calculated using the reproducibility standard deviation, calculated from the participants' results (after removal of outliers) in the relevant test round. It might occur that the z-Score between -2 and 2 covers a large range of measurement values when the variance of the results is high. On the other hand, the range of good results can be very narrow, when the variation of the participants' results is small.

The recovery rate is calculated for the individual result based on the target value and is thus independent of the reproducibility standard deviation. In the case of a high variance of the results, participants should also consider recovery rates as additional criteria to decide on the necessity of internal quality assurance measures.

This is particularly recommended for the parameters trans-1,2-Dichloroethene and n-Decane.

5 Annotations on tables and charts


5.1 Information and abbreviations in tables

Parameter	Analyte identifier
Sample	Sample identifier
Unit	Given unit for result and uncertainty (e.g. µg/l)
Mean	Mean of the participants results, without outliers (3 significant digits)
CI (99 %)	99 % confidence interval (3 significant digits)
Minimum	Minimum of all submitted results, after removal of outliers (3 significant digits)
Maximum	Maximum of all submitted results, after removal of outliers (3 significant digits)
SD	Reproducibility standard deviation, calculated from the participants results, after removal of outliers (3 significant digits)
RSD %	Reproducibility standard deviation, calculated from the participants results relative to the target value, given in %, after removal of outliers (2 significant digits)
Control test value ± U	Mean of control test value ± measurement uncertainty (3 significant digits)
Labcode	Laboratory identifier (anonymized)
Result	Result as indicated by participant (max. 5 decimal places)
± U	Results uncertainty as indicated by participant (max. 5 decimal places)
LOQ	Limit of quantification
LOD	Limit of detection
Recovery	Recovery rate in % based on target value (3 significant digits, max. one decimal place given)
z-Score	Deviation of result based on target value depicted as a multiple of the criteria (3 significant digits, max. 2 decimal places given)
-	<i>No data available</i>
Comments	Comment on the respective result (e.g. H, FN, FP)
H	Outlier according to Hampel-Test
FN	False negative – for a result < LOQ or result < LOD: The absolute value of the LOQ or LOD fulfils the condition of an outlier according to the Hampel test.
FP	False positive – for parameters where no target value is available because of a too low analyte content (n < 6):


	Result that exceeds the median of the absolute values of the transmitted LOQs or LODs by more than 100 %.
Standard deviation	Reproducibility standard deviation, calculated from the participants results (3 significant digits)
Rel. standard deviation	Reproducibility standard deviation, calculated from the participants results relative to the target value, given in %, (3 significant digits)
n	Number of results
Target value	Mean of the participants results, without outliers (3 significant digits), unless stated otherwise in section 4
Criteria	Criteria for z-Score calculation. The given value matches the reproducibility standard deviation, calculated from the participants' results, after removal of outliers (3 significant digits), unless stated otherwise in section 4

5.2 Graphical presentation of results


Example chart: Results




Example chart: Recovery



Example chart: z-score



Example chart: z-score - laboratory oriented report



Summary of results, after removal of outliers: CHC and BTEX & C5-C10 - CBL04

6 Summary of results, after removal of outliers

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Benzene	BL06 - BTEX & C5-C10	µg/tube	17	2	4.8	± 0.348	3.8	5.5	0.478	9.9
Ethylbenzene	BL06 - BTEX & C5-C10	µg/tube	19	0	5.66	± 0.659	4.27	8.16	0.958	17
o-Xylene	BL06 - BTEX & C5-C10	µg/tube	18	1	5.12	± 0.501	3.23	6.37	0.708	14
Sum of m-Xylene and p-Xylene	BL06 - BTEX & C5-C10	µg/tube	18	1	10.8	± 1.12	6.96	13.2	1.58	15
Toluene	BL06 - BTEX & C5-C10	µg/tube	18	1	5.2	± 0.407	3.95	6.1	0.576	11
n-Decane	BL06 - BTEX & C5-C10	µg/tube	11	1	4.56	± 1.28	2.4	7.29	1.41	31
n-Heptane	BL06 - BTEX & C5-C10	µg/tube	10	2	5.57	± 1.15	2.78	6.95	1.21	22
n-Hexane	BL06 - BTEX & C5-C10	µg/tube	7	3	6.02	± 0.613	4.9	6.5	0.541	9
n-Nonane	BL06 - BTEX & C5-C10	µg/tube	11	1	5.75	± 1.44	3.58	9.38	1.59	28
n-Octane	BL06 - BTEX & C5-C10	µg/tube	10	2	6.15	± 0.998	4	7.4	1.05	17
n-Pentane	BL06 - BTEX & C5-C10	µg/tube	11	1	6.2	± 0.97	4.07	7.68	1.07	17
1,1,1-Trichloroethane	CL05 - CHC	µg/tube	16	2	10.9	± 0.689	9.43	12.5	0.919	8.5
cis-1,2-Dichloroethene	CL05 - CHC	µg/tube	16	2	7.23	± 1.55	2.3	11.1	2.07	29
Tetrachloroethene	CL05 - CHC	µg/tube	17	1	11.5	± 1.14	8.14	14.2	1.57	14
Tetrachloromethane	CL05 - CHC	µg/tube	16	2	13.3	± 1.21	11.1	16.8	1.61	12
trans-1,2-Dichloroethene	CL05 - CHC	µg/tube	18	0	6.31	± 2.52	0.6	13.2	3.56	56
Trichloroethene	CL05 - CHC	µg/tube	17	1	10.1	± 1.15	6.34	13.8	1.58	16
Trichloromethane	CL05 - CHC	µg/tube	16	2	9.71	± 0.6	8.35	11.3	0.8	8.2

7 Parameter oriented report

Benzene	13
Ethylbenzene.....	17
o-Xylene	21
Sum of m- and p-Xylene.....	25
Toluene.....	29
n-Decane.....	33
n-Heptane.....	37
n-Hexane	41
n-Nonane.....	45
n-Octane.....	49
n-Pentane.....	53
1,1,1-Trichloroethane.....	57
cis-1,2-Dichloroethene.....	61
Tetrachloroethene	65
Tetrachloromethane.....	69
trans-1,2-Dichloroethene	73
Trichloroethene.....	77
Trichloromethane.....	81

Parameter oriented report

BL06 - BTEX & C5-C10

Benzene

Unit	µg/tube
Mean ± CI (99%)	4.8 ± 0.348
Minimum - Maximum	3.8 - 5.5
Control test value ± U	4.06 ± 0.349


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	4.61	0.5	96	-0.4	
LC0002	3.8	0.76	79.1	-2.1	
LC0003	3.9	0.78	81.2	-1.89	
LC0004	4.9	2	102	0.2	
LC0005	5.28	0.79	110	1	
LC0006	5.1	1.5	106	0.62	
LC0007	4.49	0.138	93.5	-0.65	
LC0009	7.6	1.52	158	5.86	H
LC0010	4.27	0.1	88.9	-1.11	
LC0011	4.69	0.94	97.7	-0.23	
LC0012	4.817	1.2	100	0.03	
LC0014	5.5	0.6	115	1.46	
LC0015	5.27	1.1	110	0.98	
LC0016	5.032	0.302	105	0.48	
LC0017	4.7	0.5	97.9	-0.21	
LC0018	5.17	0.52	108	0.77	
LC0020	4.9	1.2	102	0.2	
LC0021	5.203	1.04	108	0.84	
LC0022	3.26	0.65	67.9	-3.23	H

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	4.87 ± 0.602	4.8 ± 0.348	µg/tube
Minimum	3.26	3.8	µg/tube
Maximum	7.6	5.5	µg/tube
Standard deviation	0.875	0.478	µg/tube
rel. Standard deviation	18	9.95 %	
n	19	17	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

Ethylbenzene

Unit	µg/tube
Mean ± CI (99%)	5.66 ± 0.659
Minimum - Maximum	4.27 - 8.16
Control test value ± U	5.20 ± 0.448


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	5.71	0.6	101	0.05	
LC0002	4.5	0.9	79.5	-1.21	
LC0003	5.4	1.1	95.4	-0.27	
LC0004	5.8	2.3	102	0.15	
LC0005	6.56	0.98	116	0.94	
LC0006	6.1	1.8	108	0.46	
LC0007	4.9	0.121	86.6	-0.79	
LC0009	8.16	1.63	144	2.61	
LC0010	4.27	0.1	75.5	-1.45	
LC0011	5.64	1.1	99.7	-0.02	
LC0012	5.755	1.44	102	0.1	
LC0014	5	0.5	88.4	-0.69	
LC0015	4.99	1	88.2	-0.7	
LC0016	6.211	0.559	110	0.58	
LC0017	5.3	0.5	93.7	-0.37	
LC0018	6.83	0.68	121	1.22	
LC0020	5.1	1.3	90.1	-0.58	
LC0021	6.717	1.343	119	1.11	
LC0022	4.57	0.91	80.8	-1.14	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	5.66 ± 0.659	5.66 ± 0.659	µg/tube
Minimum	4.27	4.27	µg/tube
Maximum	8.16	8.16	µg/tube
Standard deviation	0.958	0.958	µg/tube
rel. Standard deviation	16.9	16.9	%
n	19	19	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

o-Xylene

Unit	µg/tube
Mean ± CI (99%)	5.12 ± 0.501
Minimum - Maximum	3.23 - 6.37
Control test value ± U	4.65 ± 0.4


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	4.69	0.5	91.5	-0.61	
LC0002	4.5	0.9	87.8	-0.88	
LC0003	4.5	0.9	87.8	-0.88	
LC0004	5.1	2	99.5	-0.03	
LC0005	5.81	0.87	113	0.97	
LC0006	5.3	1.6	103	0.25	
LC0007	4.71	0.127	91.9	-0.58	
LC0009	8.29	1.66	162	4.47	H
LC0010	3.23	0.1	63	-2.67	
LC0011	5.46	1.1	107	0.47	
LC0012	5.547	1.39	108	0.6	
LC0014	4.7	0.5	91.7	-0.6	
LC0015	5.4	1	105	0.39	
LC0016	5.861	0.527	114	1.04	
LC0017	5.5	0.6	107	0.53	
LC0018	6.37	0.64	124	1.76	
LC0020	5.1	1.3	99.5	-0.03	
LC0021	5.726	1.145	112	0.85	
LC0022	4.73	0.95	92.3	-0.56	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	5.29 ± 0.689	5.12 ± 0.501	µg/tube
Minimum	3.23	3.23	µg/tube
Maximum	8.29	6.37	µg/tube
Standard deviation	1	0.708	µg/tube
rel. Standard deviation	18.9	13.8	%
n	19	18	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

Sum of m-Xylene and p-Xylene

Unit	µg/tube
Mean ± CI (99%)	10.8 ± 1.12
Minimum - Maximum	6.96 - 13.2
Control test value ± U	9.45 ± 0.812


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	10.7	1	98.8	-0.08	
LC0002	9.1	1.8	84	-1.1	
LC0003	9.2	1.8	84.9	-1.03	
LC0004	11.2	4.5	103	0.23	
LC0005	12.55	1.88	116	1.09	
LC0006	11.9	3.6	110	0.68	
LC0007	9.42	0.107	87	-0.89	
LC0009	16.68	3.34	154	3.7	H
LC0010	6.96	0.1	64.2	-2.45	
LC0011	10.8	2.2	99.7	-0.02	
LC0012	11.32	2.83	104	0.31	
LC0014	10.1	1	93.2	-0.46	
LC0015	12.38	1	114	0.98	
LC0016	12.123	0.97	112	0.82	
LC0017	12.3	1.2	114	0.93	
LC0018	13.24	1.3	122	1.52	
LC0020	10.6	2.6	97.8	-0.15	
LC0021	11.693	2.339	108	0.54	
LC0022	9.42	1.88	87	-0.89	

Characteristics of parameter

	all results	without outliers	Unit
Mean ± CI (99%)	11.1 ± 1.4	10.8 ± 1.12	µg/tube
Minimum	6.96	6.96	µg/tube
Maximum	16.7	13.2	µg/tube
Standard deviation	2.04	1.58	µg/tube
rel. Standard deviation	18.3	14.6	%
n	19	18	-

Graphical presentation of results


Results



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: BL06, Parameter: Sum of m-Xylene and p-Xylene


Recovery rate



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: BL06, Parameter: Sum of m-Xylene and p-Xylene

Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

Toluene

Unit	µg/tube
Mean ± CI (99%)	5.2 ± 0.407
Minimum - Maximum	3.95 - 6.1
Control test value ± U	4.52 ± 0.388


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	5.14	0.5	98.9	-0.1	
LC0002	4.2	0.84	80.8	-1.73	
LC0003	4.8	0.96	92.3	-0.69	
LC0004	5.5	2.2	106	0.52	
LC0005	5.89	0.88	113	1.2	
LC0006	5.5	1.7	106	0.52	
LC0007	4.68	0.116	90	-0.9	
LC0009	7.62	1.52	147	4.2	H
LC0010	5.01	0.1	96.4	-0.33	
LC0011	5.29	1.1	102	0.16	
LC0012	5.552	1.39	107	0.61	
LC0014	5.2	0.5	100	0.00	
LC0015	4.95	1.2	95.2	-0.43	
LC0016	5.687	0.398	109	0.85	
LC0017	5.9	0.6	113	1.22	
LC0018	5.43	0.54	104	0.4	
LC0020	4.8	1.2	92.3	-0.69	
LC0021	6.103	1.22	117	1.57	
LC0022	3.95	0.79	76	-2.17	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	5.33 ± 0.543	5.2 ± 0.407	µg/tube
Minimum	3.95	3.95	µg/tube
Maximum	7.62	6.1	µg/tube
Standard deviation	0.789	0.576	µg/tube
rel. Standard deviation	14.8	11.1	%
n	19	18	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Decane

Unit	µg/tube
Mean ± CI (99%)	4.56 ± 1.28
Minimum - Maximum	2.4 - 7.29
Control test value ± U	5.94 ± 0.891


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	4.09	0.4	89.7	-0.33	
LC0002	-	-	-	-	
LC0003	3.2	0.64	70.2	-0.96	
LC0004	2.4	0.96	52.7	-1.53	
LC0005	-	-	-	-	
LC0006	4.8	1.4	105	0.17	
LC0007	5.56	0.135	122	0.71	
LC0009	5.7	1.14	125	0.81	
LC0010	3.63	0.2	79.6	-0.66	
LC0011	4.32	0.86	94.8	-0.17	
LC0012	5.685	1.42	125	0.8	
LC0014	-	-	-	-	
LC0015	11.42	1.1	251	4.86	H
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	3.46	0.35	75.9	-0.78	
LC0020	-	-	-	-	
LC0021	7.287	1.457	160	1.93	
LC0022	-	-	-	-	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	5.13 ± 2.07	4.56 ± 1.28	µg/tube
Minimum	2.4	2.4	µg/tube
Maximum	11.4	7.29	µg/tube
Standard deviation	2.4	1.41	µg/tube
rel. Standard deviation	46.7	31	%
n	12	11	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Heptane

Unit	µg/tube
Mean ± CI (99%)	5.57 ± 1.15
Minimum - Maximum	2.78 - 6.95
Control test value ± U	5.75 ± 0.863


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	5.27	0.5	94.6	-0.25	
LC0002	-	-	-	-	
LC0003	5.2	1	93.3	-0.31	
LC0004	4.8	1.9	86.1	-0.64	
LC0005	-	-	-	-	
LC0006	5.5	1.7	98.7	-0.06	
LC0007	6.04	0.131	108	0.39	
LC0009	6.95	1.39	125	1.14	
LC0010	6.9	0.2	124	1.1	
LC0011	6.04	1.2	108	0.39	
LC0012	6.253	1.56	112	0.56	
LC0014	-	-	-	-	
LC0015	28.04	1.3	503	18.6	H
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	2.78	0.29	49.9	-2.31	
LC0020	-	-	-	-	
LC0021	13.673	2.735	245	6.69	H
LC0022	-	-	-	-	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	8.12 ± 5.87	5.57 ± 1.15	µg/tube
Minimum	2.78	2.78	µg/tube
Maximum	28	6.95	µg/tube
Standard deviation	6.78	1.21	µg/tube
rel. Standard deviation	83.5	21.7	%
n	12	10	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Hexane

Unit	µg/tube
Mean ± CI (99%)	6.02 ± 0.613
Minimum - Maximum	4.9 - 6.5
Control test value ± U	5.21 ± 0.781


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	5.89	0.6	97.9	-0.23	
LC0002	-	-	-	-	
LC0003	6.5	1.3	108	0.89	
LC0004	4.9	2	81.4	-2.07	
LC0005	-	-	-	-	
LC0006	6.4	1.9	106	0.71	
LC0007	< 0.1 (LOQ)	-	-	-	
LC0009	6.23	1.25	104	0.39	
LC0010	3.16	0.2	52.5	-5.28	H
LC0011	6.26	1.3	104	0.45	
LC0012	5.937	1.48	98.7	-0.15	
LC0014	-	-	-	-	
LC0015	4.15	1.3	69	-3.45	H
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	-	-	-	-	
LC0020	-	-	-	-	
LC0021	18.28	3.656	304	22.7	H
LC0022	-	-	-	-	

Characteristics of parameter

	all results	without outliers	Unit
Mean ± CI (99%)	6.77 ± 3.98	6.02 ± 0.613	µg/tube
Minimum	3.16	4.9	µg/tube
Maximum	18.3	6.5	µg/tube
Standard deviation	4.19	0.541	µg/tube
rel. Standard deviation	61.9	8.99	%
n	10	7	-

Graphical presentation of results

Results




Recovery rate



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: BL06, Parameter: n-Hexane

Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Nonane

Unit	µg/tube
Mean ± CI (99%)	5.75 ± 1.44
Minimum - Maximum	3.58 - 9.38
Control test value ± U	6.22 ± 0.933


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	6.26	0.6	109	0.32	
LC0002	-	-	-	-	
LC0003	4.3	0.86	74.8	-0.91	
LC0004	3.8	1.5	66.1	-1.22	
LC0005	-	-	-	-	
LC0006	5.6	1.7	97.4	-0.09	
LC0007	6.27	0.129	109	0.33	
LC0009	6.43	1.29	112	0.43	
LC0010	6.09	0.2	106	0.21	
LC0011	5.27	1.1	91.6	-0.3	
LC0012	6.279	1.57	109	0.33	
LC0014	-	-	-	-	
LC0015	9.38	1.1	163	2.28	
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	3.58	0.36	62.3	-1.36	
LC0020	-	-	-	-	
LC0021	10.68	2.136	186	3.09	H
LC0022	-	-	-	-	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	6.16 ± 1.8	5.75 ± 1.44	µg/tube
Minimum	3.58	3.58	µg/tube
Maximum	10.7	9.38	µg/tube
Standard deviation	2.08	1.59	µg/tube
rel. Standard deviation	33.8	27.7	%
n	12	11	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Octane

Unit	µg/tube
Mean ± CI (99%)	6.15 ± 0.998
Minimum - Maximum	4 - 7.4
Control test value ± U	6.24 ± 0.937


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	5.89	0.6	95.8	-0.25	
LC0002	-	-	-	-	
LC0003	4.8	0.96	78	-1.28	
LC0004	4	1.6	65	-2.04	
LC0005	-	-	-	-	
LC0006	7.4	2.2	120	1.19	
LC0007	6.37	0.127	104	0.21	
LC0009	6.59	1.32	107	0.42	
LC0010	6.93	0.2	113	0.74	
LC0011	5.85	1.2	95.1	-0.28	
LC0012	6.8	1.7	111	0.62	
LC0014	-	-	-	-	
LC0015	6.87	1	112	0.69	
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	3.47	0.35	56.4	-2.55	H
LC0020	-	-	-	-	
LC0021	12.467	2.493	203	6.01	H
LC0022	-	-	-	-	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	6.45 ± 1.95	6.15 ± 0.998	µg/tube
Minimum	3.47	4	µg/tube
Maximum	12.5	7.4	µg/tube
Standard deviation	2.26	1.05	µg/tube
rel. Standard deviation	34.9	17.1	%
n	12	10	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

BL06 - BTEX & C5-C10

n-Pentane

Unit	µg/tube
Mean ± CI (99%)	6.2 ± 0.97
Minimum - Maximum	4.07 - 7.68
Control test value ± U	4.84 ± 0.969


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	6.2	0.6	100	0.00	
LC0002	-	-	-	-	
LC0003	5.4	1.1	87.1	-0.74	
LC0004	7.4	3	119	1.12	
LC0005	-	-	-	-	
LC0006	7.4	2.2	119	1.12	
LC0007	5.87	0.126	94.7	-0.3	
LC0009	5.63	1.13	90.9	-0.53	
LC0010	6.82	0.2	110	0.58	
LC0011	6.16	1.2	99.4	-0.03	
LC0012	5.534	1.38	89.3	-0.62	
LC0014	-	-	-	-	
LC0015	4.07	1	65.7	-1.98	
LC0016	-	-	-	-	
LC0017	-	-	-	-	
LC0018	7.68	0.77	124	1.38	
LC0020	-	-	-	-	
LC0021	24.227	4.845	391	16.8	H
LC0022	-	-	-	-	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	7.7 ± 4.59	6.2 ± 0.97	µg/tube
Minimum	4.07	4.07	µg/tube
Maximum	24.2	7.68	µg/tube
Standard deviation	5.3	1.07	µg/tube
rel. Standard deviation	68.9	17.3	%
n	12	11	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

CL05 - CHC

1,1,1-Trichloroethane

Unit	µg/tube
Mean ± CI (99%)	10.9 ± 0.689
Minimum - Maximum	9.43 - 12.5
Control test value ± U	9.40 ± 1.24


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	10.2	1	93.9	-0.72	
LC0002	10.6	2.12	97.6	-0.28	
LC0003	12	2.4	110	1.24	
LC0004	12.5	5	115	1.78	
LC0006	11	3.3	101	0.15	
LC0007	1.43	0.126	13.2	-10.3	H
LC0008	10.72	0.536	98.7	-0.15	
LC0009	10.64	2.13	98	-0.24	
LC0010	9.43	0.2	86.8	-1.56	
LC0011	9.62	1.9	88.6	-1.35	
LC0012	11.62	2.91	107	0.83	
LC0013	10.507	2.101	96.7	-0.39	
LC0015	10.5	1	96.7	-0.39	
LC0017	10	1	92.1	-0.94	
LC0019	10.63	1.44	97.9	-0.25	
LC0020	11.3	2.8	104	0.48	
LC0021	16.08	3.2	148	5.68	H
LC0022	12.5	2.5	115	1.78	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	10.6 ± 1.94	10.9 ± 0.689	µg/tube
Minimum	1.43	9.43	µg/tube
Maximum	16.1	12.5	µg/tube
Standard deviation	2.74	0.919	µg/tube
rel. Standard deviation	25.8	8.46	%
n	18	16	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

CL05 - CHC

cis-1,2-Dichloroethene

Unit	µg/tube
Mean ± CI (99%)	7.23 ± 1.55
Minimum - Maximum	2.3 - 11.1
Control test value ± U	8.60 ± 1.19


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	7.13	0.7	98.6	-0.05	
LC0002	7.6	1.52	105	0.18	
LC0003	5.6	1.4	77.4	-0.79	
LC0004	8	3.2	111	0.37	
LC0006	9.4	2.8	130	1.05	
LC0007	1.06	0.126	14.7	-2.98	H
LC0008	6.37	0.313	88.1	-0.42	
LC0009	6.69	1.34	92.5	-0.26	
LC0010	4.85	0.2	67.1	-1.15	
LC0011	6.12	1.2	84.6	-0.54	
LC0012	9.316	2.33	129	1.01	
LC0013	7.122	1.424	98.5	-0.05	
LC0015	11.09	1.1	153	1.87	
LC0017	2.3	0.2	31.8	-2.38	
LC0019	7.03	0.77	97.2	-0.1	
LC0020	7.8	2	108	0.28	
LC0021	9.28	1.8	128	0.99	
LC0022	32.7	6.5	452	12.3	H

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	8.3 ± 4.63	7.23 ± 1.55	µg/tube
Minimum	1.06	2.3	µg/tube
Maximum	32.7	11.1	µg/tube
Standard deviation	6.55	2.07	µg/tube
rel. Standard deviation	78.9	28.6	%
n	18	16	-

Graphical presentation of results

Results




Recovery rate



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: CL05, Parameter: cis-1,2-Dichloroethene

Z-score



Parameter oriented report

CL05 - CHC

Tetrachloroethene

Unit	µg/tube
Mean ± CI (99%)	11.5 ± 1.14
Minimum - Maximum	8.14 - 14.2
Control test value ± U	11.0 ± 1.45


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	10.3	1	89.9	-0.74	
LC0002	11.3	2.26	98.6	-0.1	
LC0003	12	2.2	105	0.34	
LC0004	10.4	4.2	90.8	-0.68	
LC0006	11.4	3.2	99.5	-0.04	
LC0007	1.39	0.128	12.1	-6.42	H
LC0008	12.74	0.698	111	0.82	
LC0009	11.54	2.31	101	0.05	
LC0010	9.67	0.1	84.4	-1.14	
LC0011	10.8	2.2	94.3	-0.42	
LC0012	12.96	3.24	113	0.96	
LC0013	8.141	1.62	71	-2.12	
LC0015	14.23	1	124	1.77	
LC0017	10.7	1.1	93.4	-0.48	
LC0019	10.07	1.01	87.9	-0.89	
LC0020	11.7	2.9	102	0.15	
LC0021	13.64	2.7	119	1.39	
LC0022	13.2	2.6	115	1.11	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	10.9 ± 1.99	11.5 ± 1.14	µg/tube
Minimum	1.39	8.14	µg/tube
Maximum	14.2	14.2	µg/tube
Standard deviation	2.82	1.57	µg/tube
rel. Standard deviation	25.9	13.7	%
n	18	17	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

CL05 - CHC

Tetrachloromethane

Unit	µg/tube
Mean ± CI (99%)	13.3 ± 1.21
Minimum - Maximum	11.1 - 16.8
Control test value ± U	11.5 ± 1.52


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	12.3	1.2	92.8	-0.59	
LC0002	12.4	2.48	93.5	-0.53	
LC0003	14	2.8	106	0.46	
LC0004	15.8	6.3	119	1.58	
LC0006	12	3.5	90.5	-0.78	
LC0007	1.56	0.129	11.8	-7.25	H
LC0008	61.19	2.75	462	29.7	H
LC0009	12.98	2.6	97.9	-0.17	
LC0010	12.5	0.2	94.3	-0.47	
LC0011	11.1	2.2	83.7	-1.34	
LC0012	13.98	3.5	105	0.45	
LC0013	12.593	2.518	95	-0.41	
LC0015	13.8	1	104	0.34	
LC0017	12.1	1.2	91.3	-0.72	
LC0019	11.15	1.54	84.1	-1.31	
LC0020	13.4	3.3	101	0.09	
LC0021	16.81	3.4	127	2.2	
LC0022	15.2	3	115	1.2	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	15.3 ± 8.4	13.3 ± 1.21	µg/tube
Minimum	1.56	11.1	µg/tube
Maximum	61.2	16.8	µg/tube
Standard deviation	11.9	1.61	µg/tube
rel. Standard deviation	77.8	12.2	%
n	18	16	-

Graphical presentation of results

Results




Recovery rate



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: CL05, Parameter: Tetrachloromethane

Z-score



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: CL05, Parameter: trans-1,2-Dichloroethene

Parameter oriented report

CL05 - CHC

trans-1,2-Dichloroethene

Unit	µg/tube
Mean ± CI (99%)	6.31 ± 2.52
Minimum - Maximum	0.6 - 13.2
Control test value ± U	8.11 ± 1.07


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	6.87	0.7	109	0.16	
LC0002	7.8	1.56	124	0.42	
LC0003	2.7	0.68	42.8	-1.01	
LC0004	8.9	3.6	141	0.73	
LC0006	9.5	2.7	151	0.9	
LC0007	1.13	0.129	17.9	-1.45	
LC0008	3.77	0.189	59.7	-0.71	
LC0009	3.1	0.62	49.1	-0.9	
LC0010	1.86	0.2	29.5	-1.25	
LC0011	4.16	0.83	65.9	-0.6	
LC0012	9.222	2.31	146	0.82	
LC0013	6.253	1.252	99.1	-0.02	
LC0015	13.18	1.2	209	1.93	
LC0017	0.6	0.06	9.5	-1.6	
LC0019	7.24	0.78	115	0.26	
LC0020	8.7	2.2	138	0.67	
LC0021	8.2	1.6	130	0.53	
LC0022	10.4	2.1	165	1.15	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	6.31 ± 2.52	6.31 ± 2.52	µg/tube
Minimum	0.6	0.6	µg/tube
Maximum	13.2	13.2	µg/tube
Standard deviation	3.56	3.56	µg/tube
rel. Standard deviation	56.4	56.4	%
n	18	18	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

CL05 - CHC

Trichloroethene

Unit	µg/tube
Mean ± CI (99%)	10.1 ± 1.15
Minimum - Maximum	6.34 - 13.8
Control test value ± U	9.91 ± 1.31


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	9.61	1	95.4	-0.29	
LC0002	10.3	2.06	102	0.15	
LC0003	10	2	99.3	-0.04	
LC0004	10.7	4.3	106	0.4	
LC0006	9.1	2.7	90.4	-0.61	
LC0007	1.29	0.127	12.8	-5.56	H
LC0008	10.54	0.537	105	0.3	
LC0009	9.98	2	99.1	-0.06	
LC0010	6.34	0.1	63	-2.36	
LC0011	9.31	1.9	92.5	-0.48	
LC0012	11.62	2.9	115	0.98	
LC0013	8.692	1.738	86.3	-0.87	
LC0015	11.15	1	111	0.69	
LC0017	8.4	0.8	83.4	-1.06	
LC0019	10.27	1.99	102	0.13	
LC0020	10.1	2.5	100	0.02	
LC0021	13.77	2.7	137	2.35	
LC0022	11.3	2.3	112	0.78	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	9.58 ± 1.82	10.1 ± 1.15	µg/tube
Minimum	1.29	6.34	µg/tube
Maximum	13.8	13.8	µg/tube
Standard deviation	2.57	1.58	µg/tube
rel. Standard deviation	26.9	15.7	%
n	18	17	-

Graphical presentation of results


Results



Recovery rate



Z-score



Parameter oriented report

CL05 - CHC

Trichloromethane

Unit	µg/tube
Mean ± CI (99%)	9.71 ± 0.6
Minimum - Maximum	8.35 - 11.3
Control test value ± U	9.29 ± 1.23


Labcode	Result	± U	Recovery [%]	z-score	Comments
LC0001	9.5	1	97.9	-0.26	
LC0002	10	2	103	0.36	
LC0003	9.9	2	102	0.24	
LC0004	10.4	4.2	107	0.86	
LC0006	10.5	3.2	108	0.99	
LC0007	1.27	0.127	13.1	-10.5	H
LC0008	10.66	0.544	110	1.19	
LC0009	9.1	1.82	93.7	-0.76	
LC0010	9.55	0.2	98.4	-0.2	
LC0011	9.11	1.8	93.8	-0.75	
LC0012	11.26	2.81	116	1.94	
LC0013	8.345	1.669	86	-1.7	
LC0015	9.19	1.1	94.7	-0.65	
LC0017	8.5	0.9	87.6	-1.51	
LC0019	5.01	1.03	51.6	-5.87	H
LC0020	9.5	2.4	97.9	-0.26	
LC0021	10.43	2.1	107	0.9	
LC0022	9.38	1.88	96.6	-0.41	

Characteristics of parameter


	all results	without outliers	Unit
Mean ± CI (99%)	8.98 ± 1.66	9.71 ± 0.6	µg/tube
Minimum	1.27	8.35	µg/tube
Maximum	11.3	11.3	µg/tube
Standard deviation	2.34	0.8	µg/tube
rel. Standard deviation	26.1	8.24	%
n	18	16	-

Graphical presentation of results

Results




Recovery rate



Parameter oriented report CHC and BTEX & C5-C10 - CBL04

Sample: CL05, Parameter: Trichloromethane

Z-score



8 Laboratory oriented report

The laboratory oriented report is sorted by laboratory code.


The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.61	0.5	0.478	96	-0.40
Ethylbenzene	µg/tube	5.66	± 0.659	5.71	0.6	0.958	101	0.05
o-Xylene	µg/tube	5.12	± 0.501	4.69	0.5	0.708	91.5	-0.61
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	10.7	1	1.58	98.8	-0.08
Toluene	µg/tube	5.2	± 0.407	5.14	0.5	0.576	98.9	-0.10
n-Decane	µg/tube	4.56	± 1.28	4.09	0.4	1.41	89.7	-0.33
n-Heptane	µg/tube	5.57	± 1.15	5.27	0.5	1.21	94.6	-0.25
n-Hexane	µg/tube	6.02	± 0.613	5.89	0.6	0.541	97.9	-0.23
n-Nonane	µg/tube	5.75	± 1.44	6.26	0.6	1.59	109	0.32
n-Octane	µg/tube	6.15	± 0.998	5.89	0.6	1.05	95.8	-0.25
n-Pentane	µg/tube	6.2	± 0.97	6.2	0.6	1.07	100	0.00

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.2	1	0.919	93.9	-0.72
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	7.13	0.7	2.07	98.6	-0.05
Tetrachloroethene	µg/tube	11.5	± 1.14	10.3	1	1.57	89.9	-0.74
Tetrachloromethane	µg/tube	13.3	± 1.21	12.3	1.2	1.61	92.8	-0.59
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	6.87	0.7	3.56	109	0.16
Trichloroethene	µg/tube	10.1	± 1.15	9.61	1	1.58	95.4	-0.29
Trichloromethane	µg/tube	9.71	± 0.6	9.5	1	0.8	97.9	-0.26




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	3.8	0.76	0.478	79.1	-2.10
Ethylbenzene	µg/tube	5.66	± 0.659	4.5	0.9	0.958	79.5	-1.21
o-Xylene	µg/tube	5.12	± 0.501	4.5	0.9	0.708	87.8	-0.88
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	9.1	1.8	1.58	84	-1.10
Toluene	µg/tube	5.2	± 0.407	4.2	0.84	0.576	80.8	-1.73
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.6	2.12	0.919	97.6	-0.28
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	7.6	1.52	2.07	105	0.18
Tetrachloroethene	µg/tube	11.5	± 1.14	11.3	2.26	1.57	98.6	-0.10
Tetrachloromethane	µg/tube	13.3	± 1.21	12.4	2.48	1.61	93.5	-0.53
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	7.8	1.56	3.56	124	0.42
Trichloroethene	µg/tube	10.1	± 1.15	10.3	2.06	1.58	102	0.15
Trichloromethane	µg/tube	9.71	± 0.6	10	2	0.8	103	0.36




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	3.9	0.78	0.478	81.2	-1.89
Ethylbenzene	µg/tube	5.66	± 0.659	5.4	1.1	0.958	95.4	-0.27
o-Xylene	µg/tube	5.12	± 0.501	4.5	0.9	0.708	87.8	-0.88
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	9.2	1.8	1.58	84.9	-1.03
Toluene	µg/tube	5.2	± 0.407	4.8	0.96	0.576	92.3	-0.69
n-Decane	µg/tube	4.56	± 1.28	3.2	0.64	1.41	70.2	-0.96
n-Heptane	µg/tube	5.57	± 1.15	5.2	1	1.21	93.3	-0.31
n-Hexane	µg/tube	6.02	± 0.613	6.5	1.3	0.541	108	0.89
n-Nonane	µg/tube	5.75	± 1.44	4.3	0.86	1.59	74.8	-0.91
n-Octane	µg/tube	6.15	± 0.998	4.8	0.96	1.05	78	-1.28
n-Pentane	µg/tube	6.2	± 0.97	5.4	1.1	1.07	87.1	-0.74

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	12	2.4	0.919	110	1.24
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	5.6	1.4	2.07	77.4	-0.79
Tetrachloroethene	µg/tube	11.5	± 1.14	12	2.2	1.57	105	0.34
Tetrachloromethane	µg/tube	13.3	± 1.21	14	2.8	1.61	106	0.46
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	2.7	0.68	3.56	42.8	-1.01
Trichloroethene	µg/tube	10.1	± 1.15	10	2	1.58	99.3	-0.04
Trichloromethane	µg/tube	9.71	± 0.6	9.9	2	0.8	102	0.24




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.9	2	0.478	102	0.20
Ethylbenzene	µg/tube	5.66	± 0.659	5.8	2.3	0.958	102	0.15
o-Xylene	µg/tube	5.12	± 0.501	5.1	2	0.708	99.5	-0.03
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	11.2	4.5	1.58	103	0.23
Toluene	µg/tube	5.2	± 0.407	5.5	2.2	0.576	106	0.52
n-Decane	µg/tube	4.56	± 1.28	2.4	0.96	1.41	52.7	-1.53
n-Heptane	µg/tube	5.57	± 1.15	4.8	1.9	1.21	86.1	-0.64
n-Hexane	µg/tube	6.02	± 0.613	4.9	2	0.541	81.4	-2.07
n-Nonane	µg/tube	5.75	± 1.44	3.8	1.5	1.59	66.1	-1.22
n-Octane	µg/tube	6.15	± 0.998	4	1.6	1.05	65	-2.04
n-Pentane	µg/tube	6.2	± 0.97	7.4	3	1.07	119	1.12

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	12.5	5	0.919	115	1.78
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	8	3.2	2.07	111	0.37
Tetrachloroethene	µg/tube	11.5	± 1.14	10.4	4.2	1.57	90.8	-0.68
Tetrachloromethane	µg/tube	13.3	± 1.21	15.8	6.3	1.61	119	1.58
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	8.9	3.6	3.56	141	0.73
Trichloroethene	µg/tube	10.1	± 1.15	10.7	4.3	1.58	106	0.40
Trichloromethane	µg/tube	9.71	± 0.6	10.4	4.2	0.8	107	0.86



The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.28	0.79	0.478	110	1.00
Ethylbenzene	µg/tube	5.66	± 0.659	6.56	0.98	0.958	116	0.94
o-Xylene	µg/tube	5.12	± 0.501	5.81	0.87	0.708	113	0.97
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	12.55	1.88	1.58	116	1.09
Toluene	µg/tube	5.2	± 0.407	5.89	0.88	0.576	113	1.20
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.1	1.5	0.478	106	0.62
Ethylbenzene	µg/tube	5.66	± 0.659	6.1	1.8	0.958	108	0.46
o-Xylene	µg/tube	5.12	± 0.501	5.3	1.6	0.708	103	0.25
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	11.9	3.6	1.58	110	0.68
Toluene	µg/tube	5.2	± 0.407	5.5	1.7	0.576	106	0.52
n-Decane	µg/tube	4.56	± 1.28	4.8	1.4	1.41	105	0.17
n-Heptane	µg/tube	5.57	± 1.15	5.5	1.7	1.21	98.7	-0.06
n-Hexane	µg/tube	6.02	± 0.613	6.4	1.9	0.541	106	0.71
n-Nonane	µg/tube	5.75	± 1.44	5.6	1.7	1.59	97.4	-0.09
n-Octane	µg/tube	6.15	± 0.998	7.4	2.2	1.05	120	1.19
n-Pentane	µg/tube	6.2	± 0.97	7.4	2.2	1.07	119	1.12

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	11	3.3	0.919	101	0.15
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	9.4	2.8	2.07	130	1.05
Tetrachloroethene	µg/tube	11.5	± 1.14	11.4	3.2	1.57	99.5	-0.04
Tetrachloromethane	µg/tube	13.3	± 1.21	12	3.5	1.61	90.5	-0.78
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	9.5	2.7	3.56	151	0.90
Trichloroethene	µg/tube	10.1	± 1.15	9.1	2.7	1.58	90.4	-0.61
Trichloromethane	µg/tube	9.71	± 0.6	10.5	3.2	0.8	108	0.99




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.49	0.138	0.478	93.5	-0.65
Ethylbenzene	µg/tube	5.66	± 0.659	4.9	0.121	0.958	86.6	-0.79
o-Xylene	µg/tube	5.12	± 0.501	4.71	0.127	0.708	91.9	-0.58
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	9.42	0.107	1.58	87	-0.89
Toluene	µg/tube	5.2	± 0.407	4.68	0.116	0.576	90	-0.90
n-Decane	µg/tube	4.56	± 1.28	5.56	0.135	1.41	122	0.71
n-Heptane	µg/tube	5.57	± 1.15	6.04	0.131	1.21	108	0.39
n-Hexane	µg/tube	6.02	± 0.613	<0.1 (LOQ)	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	6.27	0.129	1.59	109	0.33
n-Octane	µg/tube	6.15	± 0.998	6.37	0.127	1.05	104	0.21
n-Pentane	µg/tube	6.2	± 0.97	5.87	0.126	1.07	94.7	-0.30

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	1.43	0.126	0.919	13.2	-10.30
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	1.06	0.126	2.07	14.7	-2.98
Tetrachloroethene	µg/tube	11.5	± 1.14	1.39	0.128	1.57	12.1	-6.42
Tetrachloromethane	µg/tube	13.3	± 1.21	1.56	0.129	1.61	11.8	-7.25
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	1.13	0.129	3.56	17.9	-1.45
Trichloroethene	µg/tube	10.1	± 1.15	1.29	0.127	1.58	12.8	-5.56
Trichloromethane	µg/tube	9.71	± 0.6	1.27	0.127	0.8	13.1	-10.50



The following results were achieved:

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.72	0.536	0.919	98.7	-0.15
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	6.37	0.313	2.07	88.1	-0.42
Tetrachloroethene	µg/tube	11.5	± 1.14	12.74	0.698	1.57	111	0.82
Tetrachloromethane	µg/tube	13.3	± 1.21	61.19	2.75	1.61	462	29.70
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	3.77	0.189	3.56	59.7	-0.71
Trichloroethene	µg/tube	10.1	± 1.15	10.54	0.537	1.58	105	0.30
Trichloromethane	µg/tube	9.71	± 0.6	10.66	0.544	0.8	110	1.19




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	7.6	1.52	0.478	158	5.86
Ethylbenzene	µg/tube	5.66	± 0.659	8.16	1.63	0.958	144	2.61
o-Xylene	µg/tube	5.12	± 0.501	8.29	1.66	0.708	162	4.47
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	16.68	3.34	1.58	154	3.70
Toluene	µg/tube	5.2	± 0.407	7.62	1.52	0.576	147	4.20
n-Decane	µg/tube	4.56	± 1.28	5.7	1.14	1.41	125	0.81
n-Heptane	µg/tube	5.57	± 1.15	6.95	1.39	1.21	125	1.14
n-Hexane	µg/tube	6.02	± 0.613	6.23	1.25	0.541	104	0.39
n-Nonane	µg/tube	5.75	± 1.44	6.43	1.29	1.59	112	0.43
n-Octane	µg/tube	6.15	± 0.998	6.59	1.32	1.05	107	0.42
n-Pentane	µg/tube	6.2	± 0.97	5.63	1.13	1.07	90.9	-0.53

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.64	2.13	0.919	98	-0.24
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	6.69	1.34	2.07	92.5	-0.26
Tetrachloroethene	µg/tube	11.5	± 1.14	11.54	2.31	1.57	101	0.05
Tetrachloromethane	µg/tube	13.3	± 1.21	12.98	2.6	1.61	97.9	-0.17
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	3.1	0.62	3.56	49.1	-0.90
Trichloroethene	µg/tube	10.1	± 1.15	9.98	2	1.58	99.1	-0.06
Trichloromethane	µg/tube	9.71	± 0.6	9.1	1.82	0.8	93.7	-0.76




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.27	0.1	0.478	88.9	-1.11
Ethylbenzene	µg/tube	5.66	± 0.659	4.27	0.1	0.958	75.5	-1.45
o-Xylene	µg/tube	5.12	± 0.501	3.23	0.1	0.708	63	-2.67
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	6.96	0.1	1.58	64.2	-2.45
Toluene	µg/tube	5.2	± 0.407	5.01	0.1	0.576	96.4	-0.33
n-Decane	µg/tube	4.56	± 1.28	3.63	0.2	1.41	79.6	-0.66
n-Heptane	µg/tube	5.57	± 1.15	6.9	0.2	1.21	124	1.10
n-Hexane	µg/tube	6.02	± 0.613	3.16	0.2	0.541	52.5	-5.28
n-Nonane	µg/tube	5.75	± 1.44	6.09	0.2	1.59	106	0.21
n-Octane	µg/tube	6.15	± 0.998	6.93	0.2	1.05	113	0.74
n-Pentane	µg/tube	6.2	± 0.97	6.82	0.2	1.07	110	0.58

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	9.43	0.2	0.919	86.8	-1.56
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	4.85	0.2	2.07	67.1	-1.15
Tetrachloroethene	µg/tube	11.5	± 1.14	9.67	0.1	1.57	84.4	-1.14
Tetrachloromethane	µg/tube	13.3	± 1.21	12.5	0.2	1.61	94.3	-0.47
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	1.86	0.2	3.56	29.5	-1.25
Trichloroethene	µg/tube	10.1	± 1.15	6.34	0.1	1.58	63	-2.36
Trichloromethane	µg/tube	9.71	± 0.6	9.55	0.2	0.8	98.4	-0.20




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.69	0.94	0.478	97.7	-0.23
Ethylbenzene	µg/tube	5.66	± 0.659	5.64	1.1	0.958	99.7	-0.02
o-Xylene	µg/tube	5.12	± 0.501	5.46	1.1	0.708	107	0.47
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	10.8	2.2	1.58	99.7	-0.02
Toluene	µg/tube	5.2	± 0.407	5.29	1.1	0.576	102	0.16
n-Decane	µg/tube	4.56	± 1.28	4.32	0.86	1.41	94.8	-0.17
n-Heptane	µg/tube	5.57	± 1.15	6.04	1.2	1.21	108	0.39
n-Hexane	µg/tube	6.02	± 0.613	6.26	1.3	0.541	104	0.45
n-Nonane	µg/tube	5.75	± 1.44	5.27	1.1	1.59	91.6	-0.30
n-Octane	µg/tube	6.15	± 0.998	5.85	1.2	1.05	95.1	-0.28
n-Pentane	µg/tube	6.2	± 0.97	6.16	1.2	1.07	99.4	-0.03

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	9.62	1.9	0.919	88.6	-1.35
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	6.12	1.2	2.07	84.6	-0.54
Tetrachloroethene	µg/tube	11.5	± 1.14	10.8	2.2	1.57	94.3	-0.42
Tetrachloromethane	µg/tube	13.3	± 1.21	11.1	2.2	1.61	83.7	-1.34
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	4.16	0.83	3.56	65.9	-0.60
Trichloroethene	µg/tube	10.1	± 1.15	9.31	1.9	1.58	92.5	-0.48
Trichloromethane	µg/tube	9.71	± 0.6	9.11	1.8	0.8	93.8	-0.75




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.817	1.2	0.478	100	0.03
Ethylbenzene	µg/tube	5.66	± 0.659	5.755	1.44	0.958	102	0.10
o-Xylene	µg/tube	5.12	± 0.501	5.547	1.39	0.708	108	0.60
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	11.32	2.83	1.58	104	0.31
Toluene	µg/tube	5.2	± 0.407	5.552	1.39	0.576	107	0.61
n-Decane	µg/tube	4.56	± 1.28	5.685	1.42	1.41	125	0.80
n-Heptane	µg/tube	5.57	± 1.15	6.253	1.56	1.21	112	0.56
n-Hexane	µg/tube	6.02	± 0.613	5.937	1.48	0.541	98.7	-0.15
n-Nonane	µg/tube	5.75	± 1.44	6.279	1.57	1.59	109	0.33
n-Octane	µg/tube	6.15	± 0.998	6.8	1.7	1.05	111	0.62
n-Pentane	µg/tube	6.2	± 0.97	5.534	1.38	1.07	89.3	-0.62

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	11.62	2.91	0.919	107	0.83
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	9.316	2.33	2.07	129	1.01
Tetrachloroethene	µg/tube	11.5	± 1.14	12.96	3.24	1.57	113	0.96
Tetrachloromethane	µg/tube	13.3	± 1.21	13.98	3.5	1.61	105	0.45
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	9.222	2.31	3.56	146	0.82
Trichloroethene	µg/tube	10.1	± 1.15	11.62	2.9	1.58	115	0.98
Trichloromethane	µg/tube	9.71	± 0.6	11.26	2.81	0.8	116	1.94



The following results were achieved:

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.507	2.101	0.919	96.7	-0.39
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	7.122	1.424	2.07	98.5	-0.05
Tetrachloroethene	µg/tube	11.5	± 1.14	8.141	1.62	1.57	71	-2.12
Tetrachloromethane	µg/tube	13.3	± 1.21	12.593	2.518	1.61	95	-0.41
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	6.253	1.252	3.56	99.1	-0.02
Trichloroethene	µg/tube	10.1	± 1.15	8.692	1.738	1.58	86.3	-0.87
Trichloromethane	µg/tube	9.71	± 0.6	8.345	1.669	0.8	86	-1.70



The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.5	0.6	0.478	115	1.46
Ethylbenzene	µg/tube	5.66	± 0.659	5	0.5	0.958	88.4	-0.69
o-Xylene	µg/tube	5.12	± 0.501	4.7	0.5	0.708	91.7	-0.60
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	10.1	1	1.58	93.2	-0.46
Toluene	µg/tube	5.2	± 0.407	5.2	0.5	0.576	100	0.00
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.27	1.1	0.478	110	0.98
Ethylbenzene	µg/tube	5.66	± 0.659	4.99	1	0.958	88.2	-0.70
o-Xylene	µg/tube	5.12	± 0.501	5.4	1	0.708	105	0.39
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	12.38	1	1.58	114	0.98
Toluene	µg/tube	5.2	± 0.407	4.95	1.2	0.576	95.2	-0.43
n-Decane	µg/tube	4.56	± 1.28	11.42	1.1	1.41	251	4.86
n-Heptane	µg/tube	5.57	± 1.15	28.04	1.3	1.21	503	18.60
n-Hexane	µg/tube	6.02	± 0.613	4.15	1.3	0.541	69	-3.45
n-Nonane	µg/tube	5.75	± 1.44	9.38	1.1	1.59	163	2.28
n-Octane	µg/tube	6.15	± 0.998	6.87	1	1.05	112	0.69
n-Pentane	µg/tube	6.2	± 0.97	4.07	1	1.07	65.7	-1.98

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.5	1	0.919	96.7	-0.39
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	11.09	1.1	2.07	153	1.87
Tetrachloroethene	µg/tube	11.5	± 1.14	14.23	1	1.57	124	1.77
Tetrachloromethane	µg/tube	13.3	± 1.21	13.8	1	1.61	104	0.34
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	13.18	1.2	3.56	209	1.93
Trichloroethene	µg/tube	10.1	± 1.15	11.15	1	1.58	111	0.69
Trichloromethane	µg/tube	9.71	± 0.6	9.19	1.1	0.8	94.7	-0.65



The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.032	0.302	0.478	105	0.48
Ethylbenzene	µg/tube	5.66	± 0.659	6.211	0.559	0.958	110	0.58
o-Xylene	µg/tube	5.12	± 0.501	5.861	0.527	0.708	114	1.04
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	12.123	0.97	1.58	112	0.82
Toluene	µg/tube	5.2	± 0.407	5.687	0.398	0.576	109	0.85
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.7	0.5	0.478	97.9	-0.21
Ethylbenzene	µg/tube	5.66	± 0.659	5.3	0.5	0.958	93.7	-0.37
o-Xylene	µg/tube	5.12	± 0.501	5.5	0.6	0.708	107	0.53
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	12.3	1.2	1.58	114	0.93
Toluene	µg/tube	5.2	± 0.407	5.9	0.6	0.576	113	1.22
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-

Sample: CL05


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10	1	0.919	92.1	-0.94
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	2.3	0.2	2.07	31.8	-2.38
Tetrachloroethene	µg/tube	11.5	± 1.14	10.7	1.1	1.57	93.4	-0.48
Tetrachloromethane	µg/tube	13.3	± 1.21	12.1	1.2	1.61	91.3	-0.72
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	0.6	0.06	3.56	9.51	-1.60
Trichloroethene	µg/tube	10.1	± 1.15	8.4	0.8	1.58	83.4	-1.06
Trichloromethane	µg/tube	9.71	± 0.6	8.5	0.9	0.8	87.6	-1.51



The following results were achieved:

Sample: BL06


Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.17	0.52	0.478	108	0.77
Ethylbenzene	µg/tube	5.66	± 0.659	6.83	0.68	0.958	121	1.22
o-Xylene	µg/tube	5.12	± 0.501	6.37	0.64	0.708	124	1.76
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	13.24	1.3	1.58	122	1.52
Toluene	µg/tube	5.2	± 0.407	5.43	0.54	0.576	104	0.40
n-Decane	µg/tube	4.56	± 1.28	3.46	0.35	1.41	75.9	-0.78
n-Heptane	µg/tube	5.57	± 1.15	2.78	0.29	1.21	49.9	-2.31
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	3.58	0.36	1.59	62.3	-1.36
n-Octane	µg/tube	6.15	± 0.998	3.47	0.35	1.05	56.4	-2.55
n-Pentane	µg/tube	6.2	± 0.97	7.68	0.77	1.07	124	1.38



The following results were achieved:

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	10.63	1.44	0.919	97.9	-0.25
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	7.03	0.77	2.07	97.2	-0.10
Tetrachloroethene	µg/tube	11.5	± 1.14	10.07	1.01	1.57	87.9	-0.89
Tetrachloromethane	µg/tube	13.3	± 1.21	11.15	1.54	1.61	84.1	-1.31
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	7.24	0.78	3.56	115	0.26
Trichloroethene	µg/tube	10.1	± 1.15	10.27	1.99	1.58	102	0.13
Trichloromethane	µg/tube	9.71	± 0.6	5.01	1.03	0.8	51.6	-5.87




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	4.9	1.2	0.478	102	0.20
Ethylbenzene	µg/tube	5.66	± 0.659	5.1	1.3	0.958	90.1	-0.58
o-Xylene	µg/tube	5.12	± 0.501	5.1	1.3	0.708	99.5	-0.03
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	10.6	2.6	1.58	97.8	-0.15
Toluene	µg/tube	5.2	± 0.407	4.8	1.2	0.576	92.3	-0.69
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	11.3	2.8	0.919	104	0.48
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	7.8	2	2.07	108	0.28
Tetrachloroethene	µg/tube	11.5	± 1.14	11.7	2.9	1.57	102	0.15
Tetrachloromethane	µg/tube	13.3	± 1.21	13.4	3.3	1.61	101	0.09
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	8.7	2.2	3.56	138	0.67
Trichloroethene	µg/tube	10.1	± 1.15	10.1	2.5	1.58	100	0.02
Trichloromethane	µg/tube	9.71	± 0.6	9.5	2.4	0.8	97.9	-0.26




The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	5.203	1.04	0.478	108	0.84
Ethylbenzene	µg/tube	5.66	± 0.659	6.717	1.343	0.958	119	1.11
o-Xylene	µg/tube	5.12	± 0.501	5.726	1.145	0.708	112	0.85
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	11.693	2.339	1.58	108	0.54
Toluene	µg/tube	5.2	± 0.407	6.103	1.22	0.576	117	1.57
n-Decane	µg/tube	4.56	± 1.28	7.287	1.457	1.41	160	1.93
n-Heptane	µg/tube	5.57	± 1.15	13.673	2.735	1.21	245	6.69
n-Hexane	µg/tube	6.02	± 0.613	18.28	3.656	0.541	304	22.70
n-Nonane	µg/tube	5.75	± 1.44	10.68	2.136	1.59	186	3.09
n-Octane	µg/tube	6.15	± 0.998	12.467	2.493	1.05	203	6.01
n-Pentane	µg/tube	6.2	± 0.97	24.227	4.845	1.07	391	16.80

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	16.08	3.2	0.919	148	5.68
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	9.28	1.8	2.07	128	0.99
Tetrachloroethene	µg/tube	11.5	± 1.14	13.64	2.7	1.57	119	1.39
Tetrachloromethane	µg/tube	13.3	± 1.21	16.81	3.4	1.61	127	2.20
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	8.2	1.6	3.56	130	0.53
Trichloroethene	µg/tube	10.1	± 1.15	13.77	2.7	1.58	137	2.35
Trichloromethane	µg/tube	9.71	± 0.6	10.43	2.1	0.8	107	0.90



The following results were achieved:

Sample: BL06

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
Benzene	µg/tube	4.8	± 0.348	3.26	0.65	0.478	67.9	-3.23
Ethylbenzene	µg/tube	5.66	± 0.659	4.57	0.91	0.958	80.8	-1.14
o-Xylene	µg/tube	5.12	± 0.501	4.73	0.95	0.708	92.3	-0.56
Sum of m-Xylene and p-Xylene	µg/tube	10.8	± 1.12	9.42	1.88	1.58	87	-0.89
Toluene	µg/tube	5.2	± 0.407	3.95	0.79	0.576	76	-2.17
n-Decane	µg/tube	4.56	± 1.28	-	-	1.41	-	-
n-Heptane	µg/tube	5.57	± 1.15	-	-	1.21	-	-
n-Hexane	µg/tube	6.02	± 0.613	-	-	0.541	-	-
n-Nonane	µg/tube	5.75	± 1.44	-	-	1.59	-	-
n-Octane	µg/tube	6.15	± 0.998	-	-	1.05	-	-
n-Pentane	µg/tube	6.2	± 0.97	-	-	1.07	-	-

Sample: CL05

Parameter	Unit	Target	± CI(99%)	Result	± U	Criteria	Recovery	z-score
1,1,1-Trichloroethane	µg/tube	10.9	± 0.689	12.5	2.5	0.919	115	1.78
cis-1,2-Dichloroethene	µg/tube	7.23	± 1.55	32.7	6.5	2.07	452	12.30
Tetrachloroethene	µg/tube	11.5	± 1.14	13.2	2.6	1.57	115	1.11
Tetrachloromethane	µg/tube	13.3	± 1.21	15.2	3	1.61	115	1.20
trans-1,2-Dichloroethene	µg/tube	6.31	± 2.52	10.4	2.1	3.56	165	1.15
Trichloroethene	µg/tube	10.1	± 1.15	11.3	2.3	1.58	112	0.78
Trichloromethane	µg/tube	9.71	± 0.6	9.38	1.88	0.8	96.6	-0.41

