

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

6 Summary of results, after removal of outliers

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	\pm CI (99%)	Minimum	Maximum	SD	RSD %
2-Amino-4-methoxy-6-methyl-1,3,5-triazine	PM01 A	µg/l	1	0	-	\pm	-	0.011	0.011	-
	PM01 B	µg/l	4	0	-	\pm	-	0.082	0.11	-
	PM01 C	µg/l	2	0	-	\pm	-	0.0135	0.28	-
2,4-D	PM01 A	µg/l	15	0	0.122	\pm	0.0118	0.097	0.142	0.0152
	PM01 B	µg/l	0	0	-	\pm	-	-	-	-
	PM01 C	µg/l	15	1	0.477	\pm	0.0431	0.379	0.59	0.0556
2,6-Dichlorobenzamide	PM01 A	µg/l	15	2	2.97	\pm	0.416	1.89	4	0.537
	PM01 B	µg/l	16	1	0.382	\pm	0.0481	0.288	0.52	0.0641
	PM01 C	µg/l	2	0	-	\pm	-	0.001	0.53	-
3,5,6-Trichloro-2-pyridinol	PM01 A	µg/l	5	0	-	\pm	-	0.521	0.95	-
	PM01 B	µg/l	2	0	-	\pm	-	0.055	0.108	-
	PM01 C	µg/l	5	0	-	\pm	-	0.062	0.131	-
Alachlor	PM01 A	µg/l	13	1	0.665	\pm	0.0629	0.494	0.786	0.0756
	PM01 B	µg/l	14	0	0.255	\pm	0.0425	0.142	0.375	0.053
	PM01 C	µg/l	1	0	-	\pm	-	1.51	1.51	-
Alachlor ESA	PM01 A	µg/l	0	0	-	\pm	-	-	-	-
	PM01 B	µg/l	3	2	-	\pm	-	2.86	2.89	-
	PM01 C	µg/l	5	0	-	\pm	-	0.07	0.143	-
Alachlor OA	PM01 A	µg/l	6	0	0.131	\pm	0.0231	0.11	0.16	0.0188
	PM01 B	µg/l	0	0	-	\pm	-	-	-	-
	PM01 C	µg/l	5	1	-	\pm	-	2.71	3.63	-
Aldrin	PM01 A	µg/l	2	0	-	\pm	-	0.027	0.199	-
	PM01 B	µg/l	3	0	-	\pm	-	0.006	0.5	-
	PM01 C	µg/l	1	0	-	\pm	-	0.32	0.32	-
AMPA	PM01 A	µg/l	0	0	-	\pm	-	-	-	-
	PM01 B	µg/l	11	0	0.489	\pm	0.131	0.18	0.672	0.145

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
AMPA	PM01 C	µg/l	8	0	0.0619	± 0.00957	0.05	0.081	0.00902	15
Atrazine-2-hydroxy	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	5	2	-	± -	2.28	2.69	-	-
	PM01 C	µg/l	6	1	0.253	± 0.0186	0.229	0.273	0.0152	6
Atrazine	PM01 A	µg/l	19	2	0.17	± 0.0143	0.143	0.21	0.0208	12
	PM01 B	µg/l	19	2	0.269	± 0.0194	0.238	0.325	0.0282	10
	PM01 C	µg/l	4	0	-	± -	0.003	0.112	-	-
Azoxystrobin	PM01 A	µg/l	11	0	0.103	± 0.0135	0.08	0.133	0.0149	14
	PM01 B	µg/l	8	3	0.523	± 0.028	0.5	0.568	0.0264	5
	PM01 C	µg/l	1	0	-	± -	0.003	0.003	-	-
Bentazone	PM01 A	µg/l	1	0	-	± -	0.05	0.05	-	-
	PM01 B	µg/l	16	2	0.672	± 0.106	0.383	0.97	0.141	21
	PM01 C	µg/l	15	3	0.115	± 0.0124	0.092	0.15	0.0159	14
Bromacil	PM01 A	µg/l	13	3	0.984	± 0.0981	0.774	1.24	0.118	12
	PM01 B	µg/l	16	0	0.137	± 0.0366	0.05	0.245	0.0488	36
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Metolachlor Metabolit - CGA 368208	PM01 A	µg/l	4	0	-	± -	0.93	3.73	-	-
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	4	0	-	± -	0.105	0.401	-	-
Chloridazon	PM01 A	µg/l	1	0	-	± -	0.581	0.581	-	-
	PM01 B	µg/l	17	2	0.227	± 0.0165	0.185	0.276	0.0226	9.9
	PM01 C	µg/l	17	2	0.77	± 0.0578	0.63	0.982	0.0795	10
Clopyralid	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	10	0	0.287	± 0.0999	0.19	0.528	0.105	37
	PM01 C	µg/l	10	0	0.647	± 0.187	0.348	1.07	0.197	30
Clothianidin	PM01 A	µg/l	7	1	0.39	± 0.0238	0.356	0.413	0.021	5.4
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	8	0	0.122	± 0.0154	0.101	0.147	0.0145	12
O-demethyl azoxystrobin	PM01 A	µg/l	4	0	-	± -	0.955	1.37	-	-

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
O-demethyl azoxystrobin	PM01 B	µg/l	0	0	-	±	-	-	-	-
	PM01 C	µg/l	4	0	-	±	-	0.118	0.171	-
Dimethachlor Metabolite - CGA 369873	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	6	0	0.0674	± 0.0264	0.028	0.085	0.0216	32
	PM01 C	µg/l	5	1	-	±	-	0.404	0.551	-
Dimethachlor Metabolite - CGA 373464 (free acid)	PM01 A	µg/l	3	0	-	±	-	0.076	0.175	-
	PM01 B	µg/l	4	0	-	±	-	0.11	0.631	-
	PM01 C	µg/l	0	0	-	±	-	-	-	-
Dichlorprop	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	16	0	0.121	± 0.0118	0.094	0.158	0.0158	13
	PM01 C	µg/l	16	0	0.753	± 0.0817	0.566	1	0.109	14
Desethylatrazine	PM01 A	µg/l	20	0	0.662	± 0.0635	0.491	0.845	0.0946	14
	PM01 B	µg/l	3	0	-	±	-	0.005	0.01	-
	PM01 C	µg/l	17	2	0.222	± 0.0179	0.18	0.27	0.0246	11
Desethyldeisopropylatrazine	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	4	1	-	±	-	0.058	0.092	-
	PM01 C	µg/l	6	0	0.234	± 0.101	0.1	0.333	0.0823	35
Desethylterbutylazine	PM01 A	µg/l	2	0	-	±	-	0.004	0.014	-
	PM01 B	µg/l	15	0	0.415	± 0.0408	0.303	0.515	0.0527	13
	PM01 C	µg/l	15	0	0.0977	± 0.0107	0.071	0.121	0.0138	14
Desisopropylatrazine	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	15	1	0.0746	± 0.00888	0.061	0.099	0.0115	15
	PM01 C	µg/l	14	2	0.197	± 0.0209	0.16	0.251	0.0261	13
Dicamba	PM01 A	µg/l	8	1	0.19	± 0.0281	0.155	0.233	0.0265	14
	PM01 B	µg/l	10	0	0.833	± 0.194	0.348	1.06	0.205	25
	PM01 C	µg/l	0	0	-	±	-	-	-	-
Dieldrin	PM01 A	µg/l	4	0	-	±	-	0.006	0.117	-

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Dieldrin	PM01 B	µg/l	0	0	-	±	-	-	-	-
	PM01 C	µg/l	4	0	-	±	-	0.009	0.179	-
Dimethachlor ESA - CGA 354742	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	11	0	0.282	± 0.0626	0.151	0.369	0.0692	25
Dimethachlor	PM01 C	µg/l	11	0	0.0841	± 0.0213	0.047	0.13	0.0235	28
	PM01 A	µg/l	10	1	0.93	± 0.0718	0.798	1.08	0.0757	8.1
Dimethachlor	PM01 B	µg/l	11	0	0.136	± 0.017	0.103	0.165	0.0188	14
	PM01 C	µg/l	0	0	-	±	-	-	-	-
Dimethachlor OA - CGA 50266	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	11	0	0.102	± 0.0241	0.058	0.156	0.0267	26
Diuron	PM01 C	µg/l	11	0	0.194	± 0.046	0.12	0.298	0.0509	26
	PM01 A	µg/l	20	2	0.601	± 0.0589	0.469	0.805	0.0877	15
Dimethenamide	PM01 B	µg/l	1	0	-	±	-	0.004	0.004	-
	PM01 C	µg/l	20	2	0.259	± 0.0278	0.162	0.361	0.0414	16
Dimethenamid ESA	PM01 A	µg/l	0	0	-	±	-	-	-	-
	PM01 B	µg/l	10	1	0.65	± 0.0595	0.51	0.728	0.0627	9.6
Dimethenamid OA	PM01 C	µg/l	10	1	0.195	± 0.0111	0.18	0.216	0.0117	6
	PM01 A	µg/l	9	0	0.389	± 0.0735	0.239	0.465	0.0735	19
Dimethylsulfamide	PM01 B	µg/l	7	2	0.15	± 0.0192	0.12	0.175	0.0169	11
	PM01 C	µg/l	0	0	-	±	-	-	-	-
Desphenylchloridazon	PM01 A	µg/l	6	0	0.117	± 0.0464	0.052	0.154	0.0379	32
	PM01 B	µg/l	0	0	-	±	-	-	-	-
Desphenylchloridazon	PM01 C	µg/l	5	1	-	±	-	0.806	1.08	-
	PM01 A	µg/l	0	0	-	±	-	-	-	-
Desphenylchloridazon	PM01 B	µg/l	6	0	0.353	± 0.0349	0.316	0.387	0.0285	8.1
	PM01 C	µg/l	6	0	1.04	± 0.151	0.882	1.2	0.124	12
Desphenylchloridazon	PM01 A	µg/l	10	3	0.392	± 0.025	0.347	0.441	0.0263	6.7
	PM01 B	µg/l	11	2	2.96	± 0.175	2.58	3.21	0.194	6.5
Desphenylchloridazon	PM01 C	µg/l	0	0	-	±	-	-	-	-

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %	
Ethofumesate	PM01 A	µg/l	10	5	0.176	± 0.0139	0.147	0.206	0.0147	8.3	
	PM01 B	µg/l		1	-	± -	0.108	0.108	-	-	
	PM01 C	µg/l		0	0.719	± 0.147	0.431	1.05	0.196	27	
Flufenacet	PM01 A	µg/l	10	0	0.495	± 0.0635	0.407	0.593	0.067	14	
	PM01 B	µg/l		0	0.31	± 0.0386	0.24	0.36	0.0406	13	
	PM01 C	µg/l		0	-	± -	-	-	-	-	
Flufenacet sulfonic acid	PM01 A	µg/l	0	0	-	± -	-	-	-	-	
	PM01 B	µg/l		6	0.0996	± 0.0471	0.0465	0.156	0.0385	39	
	PM01 C	µg/l		6	0.687	± 0.284	0.329	1.04	0.231	34	
Flufenacet OA	PM01 A	µg/l	0	0	-	± -	-	-	-	-	
	PM01 B	µg/l		6	0.589	± 0.256	0.238	0.826	0.209	35	
	PM01 C	µg/l		6	0.129	± 0.0559	0.0495	0.172	0.0456	35	
Glufosinate	PM01 A	µg/l	5	0	-	± -	0.047	0.081	-	-	
	PM01 B	µg/l		0	0	± -	-	-	-	-	
	PM01 C	µg/l		4	1	± -	-	0.128	0.26	-	
Glyphosate	PM01 A	µg/l	9	3	0.936	± 0.208	0.508	1.11	0.208	22	
	PM01 B	µg/l		10	1	± 0.186	0.0296	0.13	0.242	0.0312	17
	PM01 C	µg/l		0	0	± -	-	-	-	-	
Heptachlor epoxid	PM01 A	µg/l	2	0	-	± -	0.003	0.032	-	-	
	PM01 B	µg/l		3	0	± -	-	0.0108	0.106	-	
	PM01 C	µg/l		3	0	± -	-	0.005	0.082	-	
Heptachlor	PM01 A	µg/l	2	0	-	± -	0.006	0.057	-	-	
	PM01 B	µg/l		0	0	± -	-	-	-	-	
	PM01 C	µg/l		3	0	± -	-	0.002	0.237	-	
Hexazinone	PM01 A	µg/l	15	1	0.493	± 0.0501	0.347	0.607	0.0647	13	
	PM01 B	µg/l		1	0	± -	-	0.001	0.001	-	
	PM01 C	µg/l		15	0	± 0.153	0.0248	0.071	0.198	0.032	21
Imidacloprid	PM01 A	µg/l	13	0	0.0959	± 0.0122	0.077	0.128	0.0147	15	
	PM01 B	µg/l		0	0	± -	-	-	-	-	

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Imidacloprid	PM01 C	µg/l	11	2	0.478	± 0.0323	0.42	0.543	0.0357	7.5
Iodosulfuron-methyl	PM01 A	µg/l	6	1	0.353	± 0.0406	0.324	0.403	0.0332	9.4
	PM01 B	µg/l	7	0	0.138	± 0.0204	0.121	0.173	0.018	13
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Isoproturon-desmethyl	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	6	0	0.554	± 0.0951	0.452	0.677	0.0777	14
	PM01 C	µg/l	6	0	0.194	± 0.0313	0.157	0.226	0.0255	13
Isoproturon	PM01 A	µg/l	18	2	0.86	± 0.0696	0.68	1.07	0.0984	11
	PM01 B	µg/l	19	0	0.155	± 0.0115	0.125	0.196	0.0168	11
	PM01 C	µg/l	2	0	-	± -	0.131	0.2	-	-
MCPA	PM01 A	µg/l	15	1	0.19	± 0.0291	0.131	0.274	0.0375	20
	PM01 B	µg/l	15	1	0.782	± 0.128	0.557	1.11	0.165	21
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
MCPB	PM01 A	µg/l	1	0	-	± -	0.08	0.08	-	-
	PM01 B	µg/l	12	0	0.117	± 0.0102	0.101	0.141	0.0118	10
	PM01 C	µg/l	12	0	0.238	± 0.0174	0.202	0.265	0.0201	8.4
Methyldesphenylchloridazon	PM01 A	µg/l	10	3	0.0948	± 0.00448	0.0839	0.1	0.00472	5
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Mecoprop	PM01 A	µg/l	13	5	0.186	± 0.0076	0.165	0.2	0.00913	4.9
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	16	2	0.641	± 0.0496	0.506	0.77	0.0662	10
Mesosulfuron-methyl	PM01 A	µg/l	7	0	0.566	± 0.163	0.34	0.773	0.144	25
	PM01 B	µg/l	1	0	-	± -	0.22	0.22	-	-
	PM01 C	µg/l	6	1	0.105	± 0.0287	0.072	0.144	0.0234	22
Metazachlor ESA	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	10	1	2.99	± 0.436	2.42	4.11	0.459	15
	PM01 C	µg/l	11	0	0.076	± 0.0176	0.0355	0.105	0.0194	26
Metalaxyl	PM01 A	µg/l	14	1	0.257	± 0.0125	0.237	0.294	0.0156	6.1

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Metalaxyll	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	12	3	0.61	± 0.052	0.475	0.731	0.06	9.8
Metamitron	PM01 A	µg/l	1	0	-	± -	0.007	0.007	-	-
	PM01 B	µg/l	14	0	0.262	± 0.0298	0.172	0.324	0.0372	14
	PM01 C	µg/l	14	0	0.348	± 0.0377	0.29	0.431	0.047	13
Metazachlor OA	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	7	4	0.0761	± 0.00451	0.07	0.081	0.00398	5.2
Metazachlor	PM01 A	µg/l	18	0	0.869	± 0.0718	0.697	1.03	0.102	12
	PM01 B	µg/l	18	0	0.236	± 0.0174	0.189	0.283	0.0246	10
	PM01 C	µg/l	2	0	-	± -	0.001	0.17	-	-
Metolachlor	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	19	1	0.109	± 0.0102	0.078	0.131	0.0148	14
	PM01 C	µg/l	20	1	0.442	± 0.041	0.295	0.523	0.0611	14
Metolachlor ESA	PM01 A	µg/l	11	0	0.151	± 0.0442	0.0465	0.243	0.0489	32
	PM01 B	µg/l	10	1	2.86	± 0.415	2.14	3.61	0.437	15
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Metolachlor OA	PM01 A	µg/l	10	1	3.56	± 0.543	2.3	4.16	0.573	16
	PM01 B	µg/l	11	0	0.271	± 0.0358	0.202	0.333	0.0396	15
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Metribuzin-Desamino	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	4	0	-	± -	0.259	0.309	-	-
	PM01 C	µg/l	4	0	-	± -	0.509	0.652	-	-
Metribuzin	PM01 A	µg/l	15	0	0.1	± 0.016	0.058	0.134	0.0206	21
	PM01 B	µg/l	1	0	-	± -	0.022	0.022	-	-
	PM01 C	µg/l	1	0	-	± -	0.022	0.022	-	-
Metsulfuron-methyl	PM01 A	µg/l	8	1	0.439	± 0.053	0.381	0.541	0.05	11
	PM01 B	µg/l	7	2	0.0964	± 0.00999	0.081	0.109	0.00881	9.1
	PM01 C	µg/l	1	0	-	± -	0.008	0.008	-	-

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Nicosulfuron	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	9	2	0.178	± 0.0816	0.08	0.29	0.0816	46
	PM01 C	µg/l	9	1	0.785	± 0.544	0.317	2.09	0.544	69
Metolachlor Metabolit - NOA 413173	PM01 A	µg/l	5	0	-	± -	0.228	0.498	-	-
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	4	1	-	± -	3.03	3.84	-	-
Pethoxamid	PM01 A	µg/l	8	0	0.241	± 0.0433	0.161	0.293	0.0408	17
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	8	0	0.526	± 0.061	0.459	0.623	0.0575	11
Propazine-2-hydroxy	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	6	0	0.339	± 0.135	0.242	0.529	0.11	33
	PM01 C	µg/l	5	1	-	± -	0.07	0.098	-	-
Propazine	PM01 A	µg/l	12	3	0.573	± 0.0607	0.465	0.715	0.0701	12
	PM01 B	µg/l	13	2	0.153	± 0.0238	0.091	0.196	0.0287	19
	PM01 C	µg/l	2	0	-	± -	0.001	0.02	-	-
Propiconazole	PM01 A	µg/l	8	2	0.108	± 0.0098	0.0904	0.121	0.00924	8.6
	PM01 B	µg/l	1	0	-	± -	0.13	0.13	-	-
	PM01 C	µg/l	10	0	0.457	± 0.0507	0.38	0.554	0.0534	12
Simazine	PM01 A	µg/l	21	0	0.302	± 0.0328	0.197	0.391	0.0502	17
	PM01 B	µg/l	20	1	0.0975	± 0.0125	0.061	0.125	0.0186	19
	PM01 C	µg/l	2	0	-	± -	0.01	0.035	-	-
Terbutylazine-desethyl-2-hydroxy	PM01 A	µg/l	6	0	0.0934	± 0.0199	0.078	0.119	0.0162	17
	PM01 B	µg/l	3	0	-	± -	0.0123	0.089	-	-
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Terbutylazine	PM01 A	µg/l	18	2	0.672	± 0.0378	0.571	0.792	0.0534	7.9
	PM01 B	µg/l	19	1	0.177	± 0.0133	0.139	0.22	0.0193	11
	PM01 C	µg/l	1	0	-	± -	0.02	0.02	-	-
Terbutylazine-2-hydroxy	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	6	0	0.237	± 0.0519	0.19	0.287	0.0424	18

Summary of results, after removal of outliers: Pesticides in Accordance with the Drinking Water Ordinance - PM01

Parameter	Sample	Unit	Number of results for calculation	Number of outliers	Mean	± CI (99%)	Minimum	Maximum	SD	RSD %
Terbuthylazine-2-hydroxy	PM01 C	µg/l	6	0	0.0699	± 0.0105	0.056	0.082	0.00861	12
Thiacloprid	PM01 A	µg/l	10	1	0.681	± 0.0519	0.595	0.784	0.0547	8
	PM01 B	µg/l	11	0	0.248	± 0.0248	0.216	0.305	0.0275	11
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Thiamethoxam	PM01 A	µg/l	12	0	0.1	± 0.0137	0.0768	0.13	0.0158	16
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	11	0	0.325	± 0.0452	0.248	0.43	0.05	15
Thifensulfuron-methyl	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	8	0	0.792	± 0.143	0.545	1	0.135	17
	PM01 C	µg/l	6	2	0.0758	± 0.00512	0.072	0.082	0.00418	5.5
Tolyfluanid	PM01 A	µg/l	3	0	-	± -	0.05	0.074	-	-
	PM01 B	µg/l	1	0	-	± -	0.05	0.05	-	-
	PM01 C	µg/l	1	0	-	± -	0.06	0.06	-	-
Tribenuron-methyl	PM01 A	µg/l	3	2	-	± -	0.229	0.242	-	-
	PM01 B	µg/l	0	0	-	± -	-	-	-	-
	PM01 C	µg/l	5	0	-	± -	0.29	1.49	-	-
Triclopyr	PM01 A	µg/l	8	0	0.234	± 0.0388	0.164	0.27	0.0366	16
	PM01 B	µg/l	7	1	0.588	± 0.0467	0.519	0.645	0.0412	7
	PM01 C	µg/l	0	0	-	± -	-	-	-	-
Triflusulfuron-methyl	PM01 A	µg/l	0	0	-	± -	-	-	-	-
	PM01 B	µg/l	2	0	-	± -	0.009	0.053	-	-
	PM01 C	µg/l	2	0	-	± -	0.009	0.0525	-	-
Tritosulfuron	PM01 A	µg/l	6	1	0.285	± 0.0302	0.25	0.311	0.0246	8.6
	PM01 B	µg/l	1	0	-	± -	0.23	0.23	-	-
	PM01 C	µg/l	5	1	-	± -	0.078	0.115	-	-