

Table of assigned values Sum parameters SP06

1<sup>st</sup> Edition, 17.06.2021

## Table of assigned values

Parameter	Sample	Unit	Assigned value ±	U (k=2)	Criterion	Criterion [%]
HC-Index	SP06 A - HC -Index	mg/l	0.188 ±	0.0415	0.0789	42
	SP06 B - HC -Index	mg/l	0.926 ±	0.225	0.389	42
Phenolindex	SP06 A - Phenolindex	mg/l	0.0415 ±	0.00292	0.00457	11
	SP06 B - Phenolindex	mg/l	0.367 ±	0.00493	0.0404	11

**Legend:**

- Assigned value Target value for proficiency assessment of the participants (3 significant digits)
- U (k=2) Expanded uncertainty (k=2) of the assigned value (3 significant digits)
- Criterion Specified value for the determination of the z-score in the given unit (3 significant digits)
- Criterion [%] Specified value for the determination of the z-score in % of the assigned value (2 significant digits)