

Table of assigned values Sum parameters SP07

1<sup>st</sup> Edition, 15.06.2022

## Table of assigned values

Parameter	Sample	Unit	Assigned value	±	U (k=2)	Criterion	Criterion [%]
HC-Index	SP07 A - HC-Index	mg/l	0.145	±	0.0206	0.061	42
	SP07 B - HC-Index	mg/l	1.11	±	0.112	0.465	42
Phenolindex	SP07 A - Phenolindex	mg/l	0.0702	±	0.00204	0.00772	11
	SP07 B - Phenolindex	mg/l	0.669	±	0.025	0.0736	11

**Legend:**

- Assigned value Target value for proficiency assessment of the participants (3 significant digits)  
U (k=2) Expanded uncertainty (k=2) of the assigned value (3 significant digits)  
Criterion Specified value for the determination of the z-score in the given unit (3 significant digits)  
Criterion [%] Specified value for the determination of the z-score in % of the assigned value (2 significant digits)