

Table of assigned values Sum parameters SP08

1<sup>st</sup> Edition, 15.06.2023

## Table of assigned values

Parameter	Sample	Unit	Assigned value	±	U (k=2)	Criterion	Criterion [%]
HC-Index	SP08 A - HC-Index	mg/l	0.208	±	0.0423	0.0977	47
	SP08 B - HC-Index	mg/l	1.44	±	0.246	0.604	42
Phenol index	SP08 A - Phenol index	mg/l	0.0532	±	0.00279	0.00585	11
	SP08 B - Phenol index	mg/l	0.707	±	0.0255	0.0778	11

**Legend:**

- Assigned value Target value for proficiency assessment of the participants (3 significant digits)  
U (k=2) Expanded uncertainty (k=2) of the assigned value (3 significant digits)  
Criterion Specified value for the determination of the z-score in the given unit (3 significant digits)  
Criterion [%] Specified value for the determination of the z-score in % of the assigned value (2 significant digits)