

Table of assigned values

Parameter	Sample	Unit	Assigned value ±	U (k=2)	Criterion	Criterion [%]
HC-Index	SP10 A - HC-Index	mg/l	0.122 ±	0.0284	0.049	40
	SP10 B - HC-Index	mg/l	0.856 ±	0.142	0.342	40
Phenol index	SP10 A - Phenol index	mg/l	0.0805 ±	0.00574	0.00886	11
	SP10 B - Phenol index	mg/l	0.481 ±	0.0408	0.0529	11

Legend:

Assigned value	Target value for proficiency assessment of the participants (3 significant digits)
U (k=2)	Expanded uncertainty (k=2) of the assigned value (3 significant digits)
Criterion	Specified value for the determination of the z-score in the given unit (3 significant digits)
Criterion [%]	Specified value for the determination of the z-score in % of the assigned value (2 significant digits)