

Table of assigned values

| Parameter | Sample | Unit | Assigned value ± U (k=2) | Criterion | Criterion [%] |
|--------------|------------|------|--------------------------|-----------|---------------|
| HC-Index | SP11 A KWI | mg/l | 0.21 ± 0.0348 | 0.084 | 40 |
| | SP11 B KWI | mg/l | 0.56 ± 0.104 | 0.229 | 41 |
| Phenol index | SP11 A PHI | mg/l | 0.0598 ± 0.00431 | 0.00658 | 11 |
| | SP11 B PHI | mg/l | 0.234 ± 0.029 | 0.0351 | 15 |

Legend:

| | |
|----------------|--------------------------------------------------------------------------------------------------------|
| Assigned value | Target value for proficiency assessment of the participants (3 significant digits) |
| U (k=2) | Expanded uncertainty (k=2) of the assigned value (3 significant digits) |
| Criterion | Specified value for the determination of the z-score in the given unit (3 significant digits) |
| Criterion [%] | Specified value for the determination of the z-score in % of the assigned value (2 significant digits) |